

## QUALITY CERTIFICATE Certificate of Sterilization (CoS)

**Ref:** 1011020 (Former:10601011) - Anesthesia set with probe cover

**Batch:** 2410204

**Expiry date:** 2029-10

This is certifying that the above-mentioned product has been primarily sterilized by Ethylene Oxide (EO) within a sterilization cycle development/validation performed to  $10^{-6}$  SAL in accordance with current ISO 11135 guidelines.

Sterisets Manufacturing S.A. further certifies that the above batch of production has been manufactured in compliance with Council Directive 93/42/EEC, Portuguese Law through the Decreto-Lei no. 145/2009 and the applicable requirements of EN ISO 13485:2016.

### Quality Department

*Joana Pinto*

04/12/2024

### Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)

PART OF STI GROUP

