

## QUALITY CERTIFICATE Certificate of Analysis (CoA)

**Ref:** 3010002 - Syringe Water with NaCl 0,9%, Luer Lock, 5 ml - filling 3 ml

**Batch:** 2501135

**Expiry date:** 2027-12

Tests:	Specifications:	Results
pH	4.5 – 7.0	Complies
Iron	<1 ppm	Complies
Heavy metals	<1 ppm	Complies
NaCl Concentration	0.86-0.95 g/100mL	Complies
Clarity	Transparent	Complies
Colour	Colourless	Complies
Extractable volume	≥ 3.00 mL	Complies
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

### Quality Department

*Joana Pinto*

01/04/2025

**Sterisets Manufacturing S.A.**

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)

PART OF STI GROUP

