

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3010012 - Syringe Water with NaCl 0,9%, Luer Lock, 20 ml - filling 20 ml

Batch: 2504045

Expiry date: 2028-02

| Tests: | Specifications: | Results |
|----------------------|-------------------|------------|
| pH | 4.5 – 7.0 | Complies |
| Iron | <1 ppm | Complies |
| Heavy metals | <1 ppm | Complies |
| NaCl Concentration | 0.86-0.95 g/100mL | Complies |
| Clarity | Transparent | Complies |
| Colour | Colourless | Complies |
| Extractable volume | ≥ 20.00 mL | Complies |
| Bacterial Endotoxins | < 0.5 EU/mL | <0.5 EU/mL |

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

Quality Department

Joana Pinto

07/05/2025

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP

