

## QUALITY CERTIFICATE

### Certificate of Analysis (CoA)

**Ref:** 3990125(Former:14366SNS) - Syringe NaCl 0,9% Luer Lock 10 ml (5ml fill) SNS

**Batch:** 2506026

**Expiry date:** 2028-04

Tests:	Specifications:	Results
pH	4.5 – 7.0	Complies
Iron	<1 ppm	Complies
Heavy metals	<1 ppm	Complies
NaCl Concentration	0.86-0.95 g/100mL	Complies
Clarity	Transparent	Complies
Colour	Colourless	Complies
Extractable volume	≥ 5.00 mL	Complies
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

#### Quality Department

*Joana Pinto*

01/08/2025

**Sterisets Manufacturing S.A.**

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)

PART OF STI GROUP

