

QUALITY CERTIFICATE

Certificate of Analysis (CoA)

Ref: 3990110 (Former:14363SNS) - Syringe NaCl 0,9% Luer Lock 10 ml SNS

Batch: 2508012

Expiry date: 2028-06

Tests:	Specifications:	Results
pH	4.5 – 7.0	Complies
Iron	<1 ppm	Complies
Heavy metals	<1 ppm	Complies
NaCl Concentration	0.86-0.95 g/100mL	Complies
Clarity	Transparent	Complies
Colour	Colourless	Complies
Extractable volume	≥ 10.00 mL	Complies
Bacterial Endotoxins	< 0.5 EU/mL	<0.5 EU/mL

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

Quality Department

Joana Pinto

07/10/2025

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP

