

## QUALITY CERTIFICATE

### Certificate of Analysis (CoA)

**Ref:** 3990110 (Former:14363SNS) - Syringe NaCl 0,9% Luer Lock 10 ml SNS

**Batch:** 2511009

**Expiry date:** 2028-09

| Tests:               | Specifications:   | Results    |
|----------------------|-------------------|------------|
| pH                   | 4.5 – 7.0         | Complies   |
| Iron                 | <1 ppm            | Complies   |
| Heavy metals         | <1 ppm            | Complies   |
| NaCl Concentration   | 0.86-0.95 g/100mL | Complies   |
| Clarity              | Transparent       | Complies   |
| Colour               | Colourless        | Complies   |
| Extractable volume   | ≥ 10.00 mL        | Complies   |
| Bacterial Endotoxins | < 0.5 EU/mL       | <0.5 EU/mL |

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

#### Quality Department

*Joana Pinto*

23/01/2026

**Sterisets Manufacturing S.A.**

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

[www.sterisets.com](http://www.sterisets.com)

PART OF STI GROUP

