Sterisets
Medical Products

C047/26 REV 0
IMP 365/2

# QUALITY CERTIFICATE

## Certificate of Analysis (CoA)

**Ref:** 3020010 - Syringe NaCl 0,9% Luer Lock 10 ml *S
**Batch:** 2601205
**Expiry date:** 2028-11

| Tests: | Specifications: | Results |
|---|---|---|
| pH | 4.5 – 7.0 | Complies |
| Iron | <1 ppm | Complies |
| Heavy metals | <1 ppm | Complies |
| NaCl Concentration | 0.86-0.95 g/100mL | Complies |
| Clarity | Transparent | Complies |
| Colour | Colourless | Complies |
| Extractable volume | ≥ 10.00 mL | Complies |
| Bacterial Endotoxins | < 0.5 EU/mL | <0.5 EU/mL |

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

**Quality Department**

Joana Pinto

01/04/2026

**Sterisets Manufacturing S.A.**
Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL
Phone: +31 412 236 606
**www.sterisets.com**

PART OF STI GROUP
Sti
Sterisets International