

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3040002 - Syringe Water, Luer, 10 ml - Bulk pack *S

Batch: 2604010

Expiry date: 2028-09

| Tests: | Specifications: | Results |
|---------------------------------------|-----------------|----------|
| Conductivity (20°C) | ≤ 4,3 µS/cm | Complies |
| Identification: Acidity or alkalinity | Pass test | Complies |
| Identification: Oxidizable substances | Pass test | Complies |
| Identification: Chlorides | Pass test | Complies |
| Identification: Sulfates | Pass test | Complies |
| Concentration: Ammonium | ≤ 0,2 ppm | Complies |
| Identification: Calcium and magnesium | Pass test | Complies |
| Clarity | Transparent | Complies |
| Colour | Colourless | Complies |
| Extractable volume | ≥ 10.00 mL | Complies |
| Bacterial Endotoxins | < 4.5 EU/mL | Complies |

The above mentioned product prefilled syringe Water, has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

Quality Department

Joana Pinto

19/05/2026

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP

