

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3040008 - Syringe Water with Glycerine 10%, Luer, 10 ml - Bulk pack *S

Batch: 2506061

Expiry date: 2028-03

Tests:	Specifications:	Results
pH	4.5 – 7.0	Complies
Concentration Glycerol	9.50-10.50 g/100mL	Complies
Clarity	Transparent	Complies
Colour	Colourless	Complies
Extractable volume	≥ 10.00 mL	Complies
Bacterial Endotoxins	< 4.5 EU/mL	Complies

The above mentioned product prefilled syringe Water with Glycerine 10%, has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

Quality Department

Joana Pinto

19/05/2026

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP

