

QUALITY CERTIFICATE Certificate of Analysis (CoA)

Ref: 3990220 (Former: 14364S)- Syringe NaCl 0,9% Luer Lock 20 ml

Batch: 2604198

Expiry date: 2028-12

| Tests: | Specifications: | Results |
|----------------------|-------------------|----------|
| pH | 4.5 – 7.0 | Complies |
| Iron | <1 ppm | Complies |
| Heavy metals | <1 ppm | Complies |
| NaCl Concentration | 0.86-0.95 g/100mL | Complies |
| Clarity | Transparent | Complies |
| Colour | Colourless | Complies |
| Extractable volume | ≥ 20.00 mL | Complies |
| Bacterial Endotoxins | < 0.5 EU/mL | Complies |

The above mentioned product prefilled syringe NaCl 0,9% has been manufactured by Sterisets Manufacturing S.A. and complies with the specifications for batch release according to procedure PQ 15.

Quality Department

Sónia Sousa

16/06/2026

Sterisets Manufacturing S.A.

Zona Industrial 1 – Lote 34 a 35, 4560-164 Guilhufe, Penafiel PORTUGAL

Phone: +31 412 236 606

www.sterisets.com

PART OF STI GROUP

